

RHE20/40 Setting Instructions



Customer:

Customer Purchase Order No.:

RHE transmitter will be used with RHM (please enter Product Code of Sensor)

or RHM- (please enter RHM Serial Number if RHE Transmitter will be used with existing RHM Sensor)

Please note that only I/O options which are in RHE20/40 delivery scope will be served!

Current Output 1 4mA \triangle 20mA \triangle Unit:
Fire State: [i](#)

Current Output 2 4mA \triangle 20mA \triangle Unit:
Fire State:

Digital Output 1 Measurement of [i](#)
Pulse: A mass/volume of corresponds to pulse(s).
or
Frequency: A mass/volumetric flow or density of equals to Hz.



Digital Output 2 Measurement of
Pulse: A mass/volume of corresponds to pulse(s).
or
Frequency: A mass/volumetric flow or density of equals to Hz.

Digital Status 1 Function (not applicable for Error Status) [i](#)
Lower Alarm Val. Upper Alarm Val.

Digital Status 2 Function (not appl. for Error/Direction Status)
Lower Alarm Val. Upper Alarm Val.

Digital Input 1 Function [i](#)

Digital Input 2 Function

Batch Mode	Configuration	<input type="text"/>		
	Batch Target Mass / Volume	<input type="text"/>	<input type="text"/>	
Please see information sheet for additionally required digital status output and digital input settings 				
RS485 Interface	Modbus Slave Address	<input type="text"/>	Baud Rate	<input type="text"/>
	Data Bits	8	Stop Bits	<input type="text"/>
HART	Polling ID	<input type="text"/>	Loop State	<input type="text"/>
			Preambles	<input type="text"/>
Display/HMI	Default Screen	<input type="text"/>	Display Mode	<input type="text"/>
	Backlight LED Timeout	<input type="text"/>	<input type="text"/>	seconds

To be filled by Rheonik:

RHM- _____

RHE- _____

Date: _____

Processed and Checked by: _____