



1 Best protection against electromagnetic interference is achieved by connection of the cable screens to PE.
For special cases with longer cables and small differences in earth potential between RHM and RHE installation the screens can be connected to PE_C. In case of substantial potential differences a potential equalization cable is mandatory.

- ⚠ When connecting the sensor and transmitter the respective local code for electrical installations must be considered.
Beware that there might be additional considerations for installation in hazardous areas.
- 🔌 Do not open the RHE42 under power.

Terminal assignment channel 2, drawing shows I/O option i2

Version	CH2+	CH2-
	Terminal	Terminal
i1	42	43
i2	46	47
FF, F2	77	78
PA, P2	77	78

⚠ NOTE:
Intrinsically safe I/O circuits must only be operated with certified barriers or isolating amplifiers

created / revised	
Date	05.04.2024
Drawn	H.G.Rudolph
Appr.	M.Weber



Wiring Diagram RHE42 (is)
RHE42_****_*i1_****_***

Project	
Customer	
Draw. - Rev.	E42W-E_i1_Rev_1_3
Sheet	1 / 1